

BID EVALUATION SHEET
Package No/ Tender No. 01
Name of EquipLapment and Quantity: ICE LINED VACCINE REFRIGERATOR (QTY:128)

PART -I
KNOCK DOWN CRITERIA-(COMMERCIAL EVALUATION)
 (All evaluation parameters defined below are mandatory for compliance)

Sr#	Evaluation Parameter	M/s Sind Medical Store	M/s Capri Medical	M/s SAKUF TRADING	M/s Al Makkah Enterprises
1	Original Receipt of Tender Purchase	Yes	Yes	Yes	Yes
2	Copy of CNIC of owner/CEO/MD	Yes	Yes	Yes	Yes
3	Active Tax Payer (NTN & GST)	Yes	Yes	Yes	Yes
4	Sales Tax / Income Tax Returns for last three years	Yes	No	Yes	Yes
5	Original Manufacturer/Authorized Sole Agent	Not from the manufacturer	Yes	Yes	Yes
6	Bid Security (2% of Estimated Price)	Yes	No	Yes	No
7	Complete Item with accessories and optional	Yes	Yes	Yes	Yes
8	Certificate from the Manufacturer regarding after sales services and Installation	No	Yes	Yes	No
9	Certificate from the Manufacturer regarding Installation	No	Yes	Yes	No
10	Technical Data Sheet / Brochure	Yes	Yes	Yes	Yes
11	Undertaking on notarized Stamp paper for not being blacklisted as mentioned in the bidding document	Yes	Yes	Yes	Yes
12	Undertaking on notarized Stamp paper for Brand Name, Model as mentioned in the bidding document	Yes	No	Yes	No
13	Undertaking on notarized Stamp paper regarding compliance with terms and conditions of the bidding document	Yes	Yes	Yes	Yes
14	Bid Validity	Yes	Yes	Yes	Yes
15	Delivery Period	Not as per tender requirement (90 Days DDP instead of 60 Days DDP)	Yes	Yes	Yes
16	Compliance of Warranty as er Tender	Yes	One Year instead of required Two Years	Yes	Yes
Remarks		Not-Eligible	Not-Eligible	Eligible	Not-Eligible

PART-II
KNOCK DOWN CRITERIA-(PRODUCT EVALUATION)
 (All evaluation parameters defined below are mandatory for compliance)

Item Sr No	SPECIFICATION COMPLIANCE / EVALUATION PARAMETERS					
1	Name of Equipment	Brand	FRIMED	Vestfrost Solutions	B MEDICAL SYSTEMS	TEKNA LAB
		Model	FC39V/2	VLS 404A AC	TCW 3000 AC	2000 NFP-N LB Series
		Country of Manufacturer	Italy	Denmark	Luxembourg	Italy
		Country of Origin of Product/ Model No.	Italy	Sweden (Not as per WHO Approval/Prequalification)	Luxembourg	Italy
	Compliance with defined quality standards (FDA 510K, CE MDD, Jp MHLW as mentioned in the specifications)	Not for quoted model	Non-Compliant	Compliant	Non-Compliant	
	One Year Local Market Business Histoty of the product	No	Yes	Yes	Yes	
	Availibility of quoted model on current official website of the manufacturer	Yes	Yes	Yes	Yes	Quoted Model Variant not available

Specifications / Sample Evaluation:	Not WHO Prequalified/Approved No Switchable refrigerator / vaccine freezer, No Top Door Type The Quoted product is not for vaccine storage as per borchure Vaccing Sotrage Capacity, Ambient Temperature, Climate Zone Testined and performance at 43 C, Ice Pack Storage Capacity, Hold Over Time (Freezer & Refrigerator) not mentioned. No 05 Baskets for vaccine storage.	No Switchable Type Vaccine Refrigerator / Vaccine Freezer instead only refrigerator. No Hold Over Time for Freezer mentioned. ICE Pack Storage Capacity not verifiable from Borchure. External Temperature Reading not mentioned.	Technically Acceptable	Not WHO Prequalified/Approved No Top Door Type. Contracdiction between attached Technical Offer and Brochure. No Hold Over Time mentioned (Refrigerator and Freezer). Ice pack capacity not mentioned
Technical Eligibility of Product:	Not-Eligible	Not-Eligible	Eligible	Not-Eligible
Technical Eligibility of Firm:	Not-Eligible	Not-Eligible	Eligible	Not-Eligible
BID STATUS	Non-Responsive	Non-Responsive	Responsive	Non-Responsive