



PURCHASE CELL

DIRECTORATE GENERAL HEALTH SERVICES PUNJAB
24, COOPER ROAD, LAHORE



INVITATION FOR RFP

CORRIGENDUM

Reference to RFP Documents for the procurement of Medical Devices/IVD/Kits for PHCP uploaded on the website of DGHS "www.dghs.punjab.gov.pk", the specifications of HCV PCR Kits are revised and the first line of 3rd para on page No. 21 shall be read as follows:

"The contractor shall provide the system(s) with enough throughput as to perform at least 1,800 tests or more per twenty four (24) hours shifts."

Note: - All the other terms and conditions mentioned in the RFP Documents will remain the same.


Program Manager
PHCP